



Fundamental Physical Concepts And Clinical Uses Of Computed Tomography (CT)

VANSHIKA

Lecturer COPMS ADESH University, Bathinda-Punjab

Abstract

The advancement of X-ray computed tomography (CT) has been fundamentally influenced by the discovery of X-rays, the introduction of the Radon transform, and the progress in X-ray digital data acquisition systems alongside computer technology. In contrast to traditional X-ray imaging (general radiography), CT generates cross-sectional anatomical images of internal structures based on X-ray attenuation coefficients, which approximate tissue density across nearly all body regions. This article examines the key physical principles and technical features of the CT scanner, highlighting significant developments in CT technology that have led to the creation of helical, multidetector, cone beam, portable, dual-energy, and phase-contrast CT. It also discusses integrated imaging modalities such as positron-emission-tomography CT and single-photon-emission-computed-tomography CT, as well as various clinical applications, including image acquisition parameters, CT angiography, image enhancement, flexible image visualizations, volumetric and surface rendering on computer workstations, radiation treatment planning, and target localization in radiotherapy. A comprehensive understanding of CT characteristics will enhance patient care in diagnostics and radiotherapy, potentially improving image quality and optimizing exposure doses.

Keywords: - Computed Tomography, Physical Concept, Clinical uses, Technical aspects, Radiation dose.

I. Advents of Computed Tomography

The identification of X-ray radiation represents a significant advancement in scientific knowledge. This achievement is credited to Wilhelm Conrad Roentgen from Germany, who performed the initial experiments with cathode tubes (specifically the Crookes tube) on November 8, 1895. This type of radiation exhibits fluorescence properties, has the ability to sensitize photographic film, and can penetrate opaque materials. Roentgen designated this form of radiation as X-ray radiation. X-rays are emitted and can pass through three-

dimensional objects, with the resulting penetrations captured on films or detectors, producing two-dimensional images [1]. This process is commonly known as radiography. However, challenges arise due to overlapping information, as tissues with low absorption coefficients are difficult to distinguish, and the scattering of X-rays can negatively impact the quality of the images produced. The term "computed" in computed tomography (CT) refers to the processes of calculation or reconstruction, while "tomography" is derived from the Greek words "tomo," meaning "to cut" or "section," and "graphy," which translates to "to describe." The functioning of a CT scanner relies on X-ray technology, with the standard energy levels utilized in general CT typically ranging from 100 kV to 150 kV. The algorithms employed in the reconstruction of CT images are grounded in the mathematical principles of the Radon transformation, as articulated in the Radon theorem published in 1917 by the Austrian mathematician Johann Radon [2]. He also introduced the formula for the inverse transformation. According to the Radon theorem, it is feasible to reconstruct an image from its projections; specifically, an image distribution function can be derived from an infinite collection of projections obtained through rotational scanning.

In the early 1960s, Allan M. Cormack, an American physicist of South African descent, introduced a mathematical computation method that enabled the calculation of cross-sectional images of internal structures from the projection of attenuation data obtained through X-rays. This technique utilized the rotation of the X-ray source and electronic detector around a three-dimensional object to capture data at various angles. The invention of the X-ray-based computed tomography (CT) scanner occurred in 1972, credited to British engineer Godfrey N. Hounsfield. In 1979, Hounsfield and Cormack were jointly awarded the Nobel Prize in Physiology and Medicine for their contributions to the development of CT technology [3]. The initial clinical CT scanners included a head scanner, which was installed at Atkinson Morley Hospital in London, UK, in 1972, and a whole-body scanner at Georgetown University Medical Center in the USA in 1976. During the early implementation of head CT scanners, the image acquisition parameters featured a duration of 7 minutes, an image matrix of 80×80 pixels, a scan field of 25 cm, and a spatial resolution of approximately 1.3 mm (equivalent to about 4 lines per [LP]/cm)[4].

II. Fundamental Physical Concepts of Computed Tomography.

1. X-ray absorption as it passes through an object

X-ray absorption refers to the attenuation or reduction in intensity of X-rays as they pass through an object. This phenomenon occurs due to interactions between X-ray photons and the object's atoms or molecules, resulting in energy transfer and scattering. The primary mechanisms of X-ray absorption are the photoelectric effect, Compton scattering, and Rayleigh scattering. The attenuation coefficient (μ) characterizes the absorption, depending on factors such as density, atomic number, energy, and thickness of the material. The Beer-Lambert Law describes the exponential decay of X-ray intensity (I) as it traverses the object: $I(x) = I_0 \cdot e^{(-\mu x)}$, where I_0 is the initial intensity, μ is the attenuation coefficient, and x is the object's thickness. Understanding X-ray absorption is crucial in medical imaging (CT scans, radiography), non-destructive

testing, security screening, and materials analysis [3, 4]. Different materials exhibit varying absorption characteristics. For instance, bone and metal absorb more X-rays than soft tissue or water. The energy range of X-rays also influences absorption, with lower energy X-rays being absorbed more readily. Diagnostic X-rays typically range from 30-150 KeV, while therapeutic and industrial applications use higher energy ranges.

2. Acquisition of image data through computed tomography

Computed Tomography (CT) scans acquire image data through a process involving X-ray transmission and detection. During a CT scan, an X-ray source and detector rotate around the patient's body, capturing multiple projections from different angles. The X-ray beam passes through the body, and the detectors measure the attenuated signal. This process is repeated multiple times, generating a large number of projections. The acquired projection data is then reconstructed using complex algorithms, such as filtered back projection (FBP) or iterative reconstruction [4]. These algorithms convert the projection data into cross-sectional images, representing the internal structures of the body. The reconstruction process involves: Data processing and correction for artifacts

Application of reconstruction filters

- Conversion of attenuation coefficients to Hounsfield units (HU)
- Generation of 2D or 3D images
- CT scanners employ various acquisition protocols, including:
 - Axial scanning: scanning in a single plane
 - Helical scanning: continuous scanning in a spiral trajectory
 - Cone-beam scanning: scanning with a cone-shaped X-ray beam

These protocols enable CT scans to produce high-resolution images of internal structures, facilitating diagnosis and treatment in various medical fields.

3. Reconstruction of images obtained through computed tomography

Reconstruction of images in computed tomography (CT) involves converting projection data obtained from X-ray measurements into cross-sectional images representing internal body structures. This complex process utilizes algorithms to transform attenuated X-ray signals into spatially accurate images [5].

The reconstruction process consists of:

- Data processing and correction for artifacts
- Application of reconstruction filters (e.g., Ram-Lak, Shepp-Logan)
- Conversion of attenuation coefficients to Hounsfield units (HU)
- Generation of 2D or 3D images

Key reconstruction algorithms include:

- Filtered Back projection (FBP): most commonly used
- Iterative Reconstruction (IR): reduces artifacts and noise
- Model-Based Reconstruction (MBR): incorporates system modeling
- Deep Learning-Based Reconstruction (DLR): utilizes neural networks

Reconstruction techniques enhance image quality by:

- Reducing noise and artifacts
- Improving spatial resolution
- Enhancing contrast resolution
- Enabling 3D visualization

CT reconstruction modes include:

- Axial reconstruction: single-plane imaging
- Helical reconstruction: spiral trajectory imaging
- Cone-beam reconstruction: volumetric imaging

Advances in reconstruction algorithms and techniques continue to improve CT image quality, diagnostic accuracy, and radiation dose efficiency.

4. Computed tomography numbers/Hounsfield units

The depiction of a CT scanner is represented as a digital image. This image is composed of a square matrix of pixels, known as picture elements, with each pixel corresponding to a voxel, or volume element, of the internal structure of the object being examined, such as a patient. Since a CT slice possesses a specific thickness, each pixel effectively represents a small volume element, or voxel, calculated as the product of the pixel area and the slice thickness. The dimensions of this voxel are influenced by the matrix size, the chosen field-of-view (FOV), and the thickness of the section [5]. The values assigned to these pixels reflect the average linear attenuation coefficients of the internal tissues at the respective spatial locations, and the numerical pixel matrix associated with a particular spatial location is transformed into an image by mapping the pixel values to corresponding grayscale values.

The standard CT image is generated using a square matrix of 512×512 (totaling 262,144 pixels) and features an image depth of 12 bits, allowing for 4,096 gray levels. However, advancements in technology have led to the development of matrices with dimensions of $1,024 \times 1,024$ or $2,048 \times 2,048$ pixels, known as ultrahigh resolution CT. CT numbers, also referred to as Hounsfield units (HU), represent the attenuation values of the tissues being imaged, normalized against the attenuation of water. The linear attenuation coefficient of water

is denoted as μ water, while μ pixel represents the linear attenuation coefficient of a specific pixel. Based on the definition of the CT number (HU), the CT numbers for air and water are defined as ranging from $-1,000$ to 0 , respectively. Additionally, CT numbers are assigned to various human organs [5, 6].

5. General categories of computed tomography scanners

Computed Tomography (CT) scanners are categorized based on their design, functionality, and application. The main categories include:

1. **Single-Slice CT Scanners:** These scanners use a single X-ray detector row, acquiring one slice per rotation. They are relatively low-cost and suitable for basic imaging.
2. **Multi-Slice CT Scanners:** Featuring multiple detector rows, these scanners capture multiple slices per rotation, increasing scan speed and resolution.
3. **Spiral (Helical) CT Scanners:** These scanners move the patient through the gantry in a spiral trajectory, enabling continuous data acquisition and faster scan times.
4. **Cone-Beam CT Scanners:** Using a cone-shaped X-ray beam, these scanners capture volumetric data, ideal for dental, orthopedic, and interventional imaging.
5. **High-Resolution CT Scanners:** Designed for precise imaging, these scanners offer enhanced spatial resolution, benefiting applications like mammography and orthopedics.
6. **Dual-Source CT Scanners:** Equipped with two X-ray sources, these scanners provide improved temporal resolution, suitable for cardiovascular and dynamic imaging.
7. **Portable CT Scanners:** Compact and lightweight, these scanners facilitate bedside imaging in intensive care units, emergency departments, and remote locations.

Each category addresses specific clinical needs, balancing factors like image quality, scan speed, and patient comfort.

6. Components of computed tomography scanner

A computed tomography (CT) scanner consists of several key components that work together to produce high-quality images. The main components include:

- The X-ray Tube produces X-rays, which penetrate the patient's body. The X-ray Detector array, typically composed of solid-state or scintillation materials, measures the attenuated X-ray signal. The Gantry houses the X-ray tube and detector, rotating around the patient to capture projection data.
- The Collimator defines the X-ray beam's shape and size, controlling the scan's spatial resolution. The Filter adjusts the X-ray beam's energy spectrum, optimizing image quality.

- The Data Acquisition System (DAS) processes the detector signals, converting them into digital data. The Reconstruction Computer uses algorithms to reconstruct images from the acquired data.
- The Patient Table positions the patient accurately, ensuring proper alignment. The Gantry Control System manages the rotation and movement of the gantry.

Additional components include:

- Power Supply: provides energy for the X-ray tube
- Cooling System: regulates temperature
- Radiation Shielding: protects operators and bystanders
- Operator Console: controls scan protocols and parameters
- Image Display: showcases reconstructed images

7. Advancements in computed tomography scanners

Computed tomography (CT) scanners have undergone significant advancements, revolutionizing medical imaging. Recent technological breakthroughs include high-resolution detectors, dual-energy CT, spectral CT, and iterative reconstruction algorithms. These innovations enhance image quality, reduce radiation doses, and expand clinical applications. Wide-coverage scanners and dual-source scanners enable faster scanning and improved cardiovascular imaging. Portable CT scanners provide bedside and emergency imaging capabilities.

Artificial intelligence (AI)-powered reconstruction optimizes image quality and radiation dose reduction. Clinical applications have expanded to include advanced cardiovascular imaging, oncology imaging, neuroimaging, and musculoskeletal imaging. Radiation dose reduction strategies include low-dose scanning, dose modulation, and scan protocol optimization.

Future developments promise further advancements, including photon-counting CT, phase-contrast CT, and CT-MR hybrid systems. These innovations will continue to improve diagnostic accuracy, treatment planning, and patient outcomes. As CT technology evolves, it remains a vital tool in modern healthcare.

III. Clinical Uses of Computed Tomography

Computed Tomography (CT) plays a vital role in modern healthcare, with diverse clinical applications. Here are some uses:

- CT scans aid in diagnosing and monitoring various conditions, including cardiovascular disease, cancer, neurological disorders, and musculoskeletal injuries. They provide detailed images of internal structures, guiding surgeons and interventional radiologists during procedures.
- In emergency medicine, CT scans quickly assess trauma, stroke, and other life-threatening conditions. They also evaluate abdominal pain, appendicitis, and bowel obstruction.

- CT angiography visualizes blood vessels, detecting vascular diseases, aneurysms, and blockages. CT urography examines the urinary tract, identifying kidney stones, tumors, and other abnormalities.
- In oncology, CT scans stage cancer, monitor treatment response, and detect recurrences. They also guide biopsies and tumor ablations.
- CT scans are essential in neuroimaging, diagnosing conditions like stroke, brain tumors, and spinal injuries. They also evaluate chronic conditions such as multiple sclerosis and Alzheimer's disease.

Additionally, CT scans are used in:

- Orthopedic imaging
- Pediatric imaging
- Dental imaging
- Radiotherapy planning
- Image-guided interventions

CT's versatility, speed, and accuracy make it an indispensable diagnostic tool in modern medicine.

IV. Clinical challenges of Computed Tomography scanner technology

Computed Tomography (CT) scanner technology presents several clinical challenges, including:

Image artifacts and noise compromise image quality, leading to diagnostic errors. Metal artifacts from implants or foreign objects can cause streaks and distortions. Motion artifacts from patient movement or respiratory issues can blur images. Beam hardening artifacts from dense materials like bone or metal can affect accuracy. Radiation exposure remains a concern, particularly for pediatric and pregnant patients. High doses increase cancer risk and genetic damage. Dose creep and unnecessary scans exacerbate this issue. Limited spatial resolution and contrast sensitivity can hinder detection of small lesions or subtle abnormalities. Partial volume averaging and slice thickness limitations can lead to inaccurate diagnoses [7, 8]. Artifact correction techniques and reconstruction algorithms can introduce additional challenges, such as increased noise or altered image texture. Patient factors, including obesity, claustrophobia, and respiratory issues, can complicate scanning and image quality. Scanner calibration, maintenance, and quality control are crucial to ensuring accurate and reliable image production. Advances in CT technology aim to address these challenges, including:

- Iterative reconstruction algorithms
- Metal artifact reduction techniques
- Dose modulation and reduction strategies
- High-resolution detectors
- Advanced artifact correction methods

Addressing these clinical challenges enhances CT's diagnostic accuracy, patient safety, and overall effectiveness.

V. Radiation dose of Computed Tomography

The amount of radiation exposure from CT scans is influenced by several factors, including the specific type of CT scanner used, the nature of the examination, the dimensions of the body part being scanned, and various image acquisition parameters such as kilovolt peak (KVp), mill ampere (mA), exposure duration, X-ray beam collimation, scan pitch, and available dose reduction techniques. For abdominal scans conducted under conditions of 120 kV, 200 mA, and a 2-second exposure time with a slice length of 100 mm and a phantom diameter of 320 mm, the radiation doses measured were 21 mGy at the center and 35 mGy at the periphery of the phantom [8]. The effective radiation doses associated with CT diagnostic procedures are generally estimated to fall within the range of 1 to 10 mSv. To minimize patient dose risks, clinicians follow:

- ALARA (As Low As Reasonably Achievable) principles
- Dose modulation and reduction strategies
- Scan protocol optimization
- Patient selection and justification

VI. Conclusions: - A comprehensive understanding of CT characteristics will enhance the effectiveness and accuracy of patient care in diagnostics and radiotherapy, potentially resulting in improved image quality and optimized radiation exposure doses.

VII. References

1. Goodman PC. The new light: discovery and introduction of the X-ray. *AJR Am J Roentgen*. 1995; 165:1041-1045.



2. Mould RF. Invited review: Rontgen and the discovery of X-rays. *Br J Radiol*. 1995; 68:1145-1176.




3. Beatty J. The radon transform and the mathematics of medical imaging. Waterville: Colby College, 2012[cited 2021 Feb 15]. Available from: <https://digitalcommons.colby.edu/cgi/viewcontent.cgi?article=1649&context=honorsthesis>.

4. Barrett HH. III the radon transform and its applications. *Prog Opt*. 1984; 21:217-286.

5. Honegger J, Maier A, Kowarschik M, CT Image reconstruction basics. *Radiology Key*, 2018 [cited 2021 Feb 15]. Available from: <https://radiologykey.com/ct-image-reconstruction-basics/>.

6. Maier A, Steidl S, Christlein V, Honegger J. Medical imaging systems: an introductory guide. Cham: Springer Nature Switzerland AG; 2018:148-167.

7. Goldman LW. Principles of CT and CT technology. *J Nucl Med Technol*. 2007; 35:115-128. Quiz 129-130.



8. Bae KT, Whiting BR. Basic principles of computed tomography physics and technical considerations. *Radiology Key*, 2016 [cited 2021 Feb 15]. Available from: <https://radiologykey.com/basic-principles-of-computed-tomographyphysics-and-technical-considerations/>.

9. Hounsfield GN. Computerized transverse axial scanning (tomography). 1. Description of system. *Br J Radiol*. 1973; 46:1016-1022.



10. Rincon-Guio C, Gómez AM, Charry JD. The role of computed tomography as a prognostic tool in traumatic brain trauma. *Imaging Med*. 2017; 9:171-178.